



Thermo Scientific
Accucore HPLC Columns
Technical Guide

Ultimate Core Performance

to Maximize Your Investment



Distributed by Greyhound Chromatography and Allied Chemicals
sales@greyhoundchrom.com

Thermo
S C I E N T I F I C

Core Enhanced Technology

To meet modern requirements for outstanding performance free from constraints on the type of instruments that can be used we have developed Accucore™ HPLC columns using an exciting new concept.

The key components of Core Enhanced Technology

Solid Core Particles

2.6µm diameter particles with a solid core generate high speed, high resolution separations without excessive backpressure

Tight Control of Particle Diameter


Enhanced selection process keeps particle size distribution to a minimum and produces high efficiency columns

Automated Packing Process

Enhanced automated procedures ensure that all columns are packed with the highest quality

Advanced Bonding Technology

Optimized phase bonding creates a series of high coverage, robust phases




	Solid Core Particles	Tight Control of Particle Diameter	Advanced Bonding Technology	Automated Packing Process
Fast Separations	•	•	•	•
High Peak Capacity	•	•	•	•
Increased Sensitivity	•	•	•	•
Lower Pressure	•	•		•
Loading Capacity		•	•	•
Reproducible Chromatography		•	•	•
Long Lifetime	•	•	•	•
Wide Selectivity			•	

The table above shows how the components of Core Enhanced Technology contribute to the outstanding features of Accucore HPLC columns.

Table of Contents

Particle Evolution	2
Theory of Fast HPLC	3
Why Core Enhanced Technology Works	4
Core Enhanced Technology Effect	5
Faster than 5 μ m and 3 μ m	6
Short Columns for Even Faster Separations	7
Higher Peak Capacity than 5 μ m and 3 μ m	8
More Sensitive than 5 μ m and 3 μ m	9
Equivalent Performance to Sub-2 μ m with Lower Pressure	10
Loading Capacity	11
Simple Method Transfer	12
UHPLC System Not Required	13
Instrument Optimization	14
Reproducible Chromatography	15
Long Lasting Columns	16
Column Protection	18
Thermo Scientific Accucore Optimum Selectivity	19
RP-MS	20
C18	21
aQ	22
Phenyl-Hexyl	23
PFP	24
HILIC	25
Applications	
Food	26
Environmental	27
Clinical	28
Forensic Toxicology	30
Forensics	30
Pharmaceuticals	31
Bio-pharmaceuticals	33
General	34
Thermo Scientific Options for Fast HLPC	35
Ordering Information	
Columns and Guards	36
Kits	37




Particle Evolution

In the search for ever faster and better separations the size and shape of column packing materials has evolved in the decades since the invention of HPLC.

Packing materials have changed from large pellicular particles via smaller totally porous particles to spherical particles with diameters of less than $2\mu\text{m}$.

Our Core Enhanced Technology has changed things again. These particles are not totally porous, but rather have a solid core and a porous outer layer.





Core Enhanced Technology

Core enhanced technology offers the advantages associated with sub- $2\mu\text{m}$ packing materials – fast, high-resolution separations – without the high backpressures required to work with the smaller particles.

Theory of Fast HPLC


The Van Deemter curves and kinetic plot below show the impact of packing material particle diameter and type on chromatographic efficiency.

The highest efficiencies and lowest rate of efficiency loss with increasing flow rate are seen with the smallest particle diameter and the Core Enhanced Technology.



Resolution

The general chromatographic resolution equation shows that resolution is directly proportional to the square root of efficiency. This means that with small diameter and core enhanced technology relatively high flow rates can be used, whilst maintaining the high efficiencies required for good separations.



α Selectivity

N Efficiency

k' Capacity factor


$$R_s = \frac{1}{4} \frac{(\alpha - 1)}{\alpha} \sqrt{N} \frac{k'}{1 + k'}$$

Why Core Enhanced Technology Works

Two factors that strongly affect chromatographic efficiency are mass transfer and eddy diffusion, the C and A terms respectively from the Van Deemter equation.


$$H = A + \frac{B}{u} + Cu$$

- H* Height equivalent to theoretical plate (column length/efficiency)
- A* Eddy diffusion
- B* Longitudinal diffusion
- C* Resistance to Mass Transfer
- u* Mobile phase linear velocity



Resistance to mass transfer is minimized by the solid core design of Core Enhanced Technology particles as the diffusional path of analytes is limited by the depth of the outer porous layer.

Material	Accucore 2.6µm
Average Particle Size	1.12
Distribution (D90/10)	



The tight control of Core Enhanced Technology particle diameter and automated packing process used for Accucore HPLC columns result in a tight, highly uniform packed bed that minimizes eddy diffusion.

Lower Backpressure

- L* Column length (cm)
- η Mobile phase viscosity (cP)
- F* Flow rate (mL/min)
- d_p^2 Particle diameter (µm)
- d_c^2 Column diameter (cm)



$$\Delta P \sim \frac{250L\eta F}{d_p^2 d_c^2}$$

The equation above shows how backpressure is related to particle diameter.

The tightly controlled 2.6µm diameter of Accucore particles also results in lower backpressures than typically seen with sub-2µm materials.


Core Enhanced Technology Effect

The plots below show how the efficiency and backpressure of Accucore HPLC columns compare to columns packed with traditional totally porous 5 μ m, 3 μ m and 1.9 μ m particles.



Faster than 5 μ m and 3 μ m

Using Accucore HPLC columns excellent separations can be achieved in shorter times. The examples on this page show how by increasing flow rates while maintaining efficiency, and therefore resolution, the time taken to separate a mixture can be reduced by a factor of 3 and solvent costs can be reduced by 7-times!




Mobile phase:	A – Water; B – Acetonitrile
Gradient:	Accucore RP-MS 2.6μm 100 x 2.1mm = 35 – 60% B in 3.5 minutes fully porous 3μm 100 x 2.1mm = 35 – 60% B in 4.0 minutes fully porous 5μm 100 x 2.1mm = 35 – 60% B in 6.7 minutes fully porous 5μm 150 x 4.6mm = 35 – 60% B in 10.0 minutes
Flow rate:	Accucore RP-MS 2.6 μ m 100 x 2.1mm = 400 μ L/min fully porous 3μm 100 x 2.1mm = 350 μ L/min fully porous 5μm 100 x 2.1mm = 210 μ /min fully porous 5μm 150 x 4.6mm = 1000 μ L/min
Injection volume:	1 μ L (fully porous 5 μ m 150 x 4.6mm = 5 μ L)
Temperature:	30°C
Detection:	UV at 247nm (0.1s rise time, 20Hz)
Analytes:	1. Tebuthiuron 2. Metoxuron 3. Monuron 4. Chlorotoluron 5. Diuron 6. Linuron

	Accucore RP-MS 2.6 μ m, 100 x 2.1mm	Fully Porous 3 μ m, 100 x 2.1mm	Fully Porous 5 μ m, 100 x 2.1mm	Fully Porous 5 μ m, 150 x 4.6mm
Resolution (critical pair)	2.50	1.96	1.64	2.64
Run time (min) including gradient re-equilibration	6.00	7.00	11.50	17.00

Reducing analysis time and solvent costs results in higher throughput and lower cost per analysis.

Short Columns for Even Faster Separations

The separating power of Accucore HPLC columns means that by using shorter column dimensions acceptable resolution can be maintained, with even greater increases in throughput and reduction in costs.



Mobile phase: A – Water; B – Acetonitrile
 Gradient: Accucore RP-MS 2.6µm
 50 x 2.1mm = 35 – 60% B in 1.8 minutes
 Accucore RP-MS 2.6µm
 100 x 2.1mm = 35 – 60% B in 3.5 minutes
 Flow rate: 400µL/min

Analysis Time and Solvent Savings

	Accucore RP-MS 2.6µm, 50 x 2.1mm	Accucore RP-MS 2.6µm, 100 x 2.1mm
Resolution (critical pair)	1.51	2.50
Run time (min) including gradient re-equilibration	3.00	6.00

A 50mm column gives acceptable separation with a doubling of productivity and halving of solvent costs.

ACCELA
600 Pump

Higher Peak Capacity than 5µm or 3µm


As an alternative to speeding up analysis the high resolution offered by Accucore HPLC columns can also be used to improve complex separations through an increase in peak capacity.

n_c Peak capacity

t_g Gradient time


W Peak width (10% height)

$$n_c = 1 + \left(\frac{t_g}{W} \right)$$



Mobile phase:	A – Water; B – Acetonitrile
Gradient:	65 – 95% B in 2.1 minutes, 95% B for 0.4 minute
Flow rate:	400µL/min
Injection volume:	1µL
Temperature:	40°C
Detection:	UV at 247nm (0.1s rise time, 20Hz)
Analytes:	1. Acetophenone 2. Propiophenone 3. Butyrophenone 4. Valerophenone 5. Hexanophenone 6. Heptanophenone 7. Octanophenone

Peak Capacity Comparison



Accucore RP-MS 2.6µm	158
fully porous 3µm	132
fully porous 5µm	100

The higher the peak capacity the more analytes can be identified within a run.



More Sensitive than 5µm or 3µm

According to the formula shown below, the sharper, taller peaks obtained with Accucore HPLC columns result in a higher signal to noise ratio (S/N) and therefore better sensitivity.

c_{max} Concentration at peak apex

N Efficiency


V_i Injection volume

L Column length

d_c Column internal diameter

k' Capacity factor

$$c_{max} \propto \frac{\sqrt{N} V_i}{L d_c^2 (1 + k')}$$



Mobile phase: A – Water; B – Acetonitrile

Gradient: Accucore RP-MS 2.6µm
100 x 2.1mm = 35 – 60% B in 3.5 minutes
fully porous 3µm
100 x 2.1mm = 35 – 60% B in 4.0 minutes
fully porous 5µm
100 x 2.1mm = 35 – 60% B in 6.7 minutes

Flow rate: Accucore RP-MS 2.6µm
100 x 2.1mm = 400µL/min
fully porous 3µm
100 x 2.1mm = 350µL/min
fully porous 5µm
100 x 2.1mm = 210µL/min

Injection volume: 1µL

Temperature: 30°C

Detection: UV at 247nm (0.1s rise time, 20Hz)

Analytes: 1. Tebuthiuron
2. Metoxuron
3. Monuron
4. Chlorotoluron
5. Diuron
6. Linuron


Sensitivity

Column	S/N (6-sigma) for Monuron	Increase in Sensitivity
Accucore 2.6µm, 100 x 2.1mm	399	136%
fully porous 3µm, 100 x 2.1mm	368	117%
fully porous 5µm, 100 x 2.1mm	169	–

Better sensitivity allows reliable detection and determination of small peaks, for example low level impurities.

Equivalent Performance to Sub-2 μ m with Lower Pressure

With solid core design, tight particle size distribution and uniform packed bed Accucore HPLC columns have broadly equivalent performance to sub-2 μ m columns and yet generate only a fraction of the backpressure.



Mobile phase:	A – Water; B – Acetonitrile
Gradient:	65 – 95% B in 1.7 minutes 95% B for 0.3 minute
Flow rate:	500 μ L/min
Injection volume:	1 μ L
Temperature:	40°C
Detection:	UV at 247nm (0.1s rise time, 20Hz)
Analytes:	1. Acetophenone 2. Propiophenone 3. Butyrophenone 4. Valerophenone 5. Hexanophenone 6. Heptanophenone 7. Octanophenone

Pressure


	Accucore RP-MS 2.6 μ m, 100 x 2.1mm	Fully Porous < 2 μ m, 100 x 2.1mm
Resolution (critical pair)	3.72	4.20
Run time (min)	3.50	3.50
Maximum pressure (bar)	171	338

Lower backpressure eliminates the requirement for UHPLC systems with maximum pressure ratings >600 bar. If a UHPLC system is used then the lower backpressure reduces wear on the instrument.

Loading Capacity


With tightly packed beds and high bonded phase coverage Accucore HPLC columns have loading capacities that allow a wide range of analyte concentrations to be determined.

The example below shows minimal change in retention and peak shape with increasing analyte concentration.



Column: Accucore RP-MS 100 x 2.1mm
 Mobile phase: 68:32 (v/v) Water/Methanol
 Flow rate: 1.0mL/min
 Temperature: 40° C
 Detection: 254nm
 Injection volume: 1µL

Concentration (ng/µL)	Load on Column (µg)
5	0.005
25	0.025
50	0.050
250	0.250
500	0.500
1000	1.000
2000	2.000



Simple Method Transfer

Fast HPLC is often performed using lower volume columns.

A few simple steps are required to transfer a method to a lower volume Accucore HPLC column.

Method Transfer Tool

A convenient method transfer tool is available at the Chromatography Resource Center www.thermoscientific.com/crc

- **Adjust Flow Rate**


Keep linear velocity constant between original and new method, taking into account particle size and geometry

- **Adjust Injection Volume**

Keep the ratio of injection volume to column volume constant

- **Adjust Gradient Profile**


Keep the number of column volumes constant for each gradient segment



Mobile phase:	A – Water; B – Acetonitrile
Gradient:	Accucore RP-MS 2.6µm 100 x 2.1mm = 35 – 60% B in 3.5 minutes fully porous 5µm 150 x 4.6mm = 35 – 60% B in 10.0 minutes
Flow rate:	Accucore RP-MS 2.6µm 100 x 2.1mm = 400µL/min fully porous 5µm 150 x 4.6mm = 1000µL/min
Injection volume:	Accucore RP-MS 2.6µm 100 x 2.1mm = 1µL fully porous 5µm 150 x 4.6mm = 5µL
Temperature:	30°C
Detection:	UV at 247nm (0.1s rise time, 20Hz)
Analytes:	1. Tebuthiuron 2. Metoxuron 3. Monuron 4. Chlorotoluron 5. Diuron 6. Linuron

UHPLC System Not Required

The low backpressures generated associated with Core Enhanced Technology mean that Accucore HPLC columns can be used with both UHPLC and HPLC systems.



Column:	Accucore RP-MS 2.6µm, 100 x 2.1mm
Mobile phase:	A – Water; B – Acetonitrile
Gradient:	65 – 95% B in 2.1 minutes 95% B for 0.4 minute
Flow rate:	400µL/min
Injection volume:	1µL
Temperature:	40°C
Detection:	UV at 247nm (0.1s rise time, 20Hz)
Analytes:	Phenones 1. Acetophenone 2. Propiophenone 3. Butyrophenone 4. Valerophenone 5. Hexanophenone 6. Heptanophenone 7. Octanophenone

System Comparison

The UHPLC system gives the best performance and any HPLC system can benefit from the faster, high resolution separations offered by Accucore HPLC columns. The higher resolution observed with the Surveyor is caused by the gradient delay.

	Accela 1250	Surveyor	Agilent 1100
Run Time (min)	2.5	3.0	3.5
Average Peak Width at 50% height (min)	0.02	0.02	0.04
Average Resolution (USP)	6.15	6.53	5.33

In order to get the best out of Accucore HPLC columns the system should be optimized for high efficiency separations.

See Instrument Optimization on page 14.



Instrument Optimization

Accucore HPLC columns produce very narrow peaks. In order to preserve this efficiency the HPLC system should be optimized to reduce any potential causes of peak broadening.

Potential causes of peak broadening are:

Extra-column band broadening

The following equation for extra-column broadening shows that it is important to limit injection volume, minimize flow cell volume and make sure that short, narrow ID tubing is used.

K Constant

V_{inj} Injection volume

V_{cell} Flow cell volume

F Flow rate

r_c Tubing radius

l_c Tubing length

D_m Diffusion coefficient in mobile phase


$$\sigma_{ext}^2 = \left(K_{inj} \frac{V_{inj}^2}{12} \right) + \left(K_{cell} \frac{V_{cell}^2}{12} + \pi^2 F^2 \right) + \left(\frac{r_c^4 l_c F}{7.6 D_m} \right)$$

Slow detector response

The detector time constant or sampling rate must be optimized for narrow peaks. If this is not done then losses in intensity and increases in peak width are seen.


	Data point*	Peak width 4 σ (s)	Peak Area	Peak height (mAu)
1Hz	2	2.04	246330	107.4
5Hz	6	0.96	57244	118.4
10Hz	10	0.87	55750	114.5
20Hz	18	0.87	55319	115.4

* Number of data points are collected over 4 σ



Fast Gradients

For fast gradients it is also important to minimize the pump dwell volume to ensure that the gradient reaches the column as quickly as possible.



Column:	fully porous < 2 μ m, 50 x 2.1mm
Mobile phase:	A – Water + 0.1% Formic Acid B – Acetonitrile + 0.1% Formic Acid
Gradient:	5 to 100% B in 2 min
Flow rate:	0.55mL/min
Temperature:	25°C
Detection:	UV@ 270nm (2 μ L flow cell)
Tubing column – detector:	0.005" ID
Injection volume:	0.5 μ L
Analytes:	1. Sulphaguandine 2. Sulphamerazine 3. Sulphamonomethoxine 4. Sulphaquinoxaline















Reproducible Chromatography

The advanced bonding technology and automated packing process used for Accucore HPLC columns result in exceptionally reproducible chromatography.

Packing Material Analysis



Every batch of Accucore packing material manufactured is analyzed, compared with specifications and issued with a certificate of analysis.

Material Properties and Analysis Based on Tanaka Tests

Silica Support Properties	Bonded Phase Properties	Hydrophobic Interactions	Secondary Interactions	Acidic Interactions	HILIC Interactions
Surface Area 	Carbon load 	Hydrophobic retention  HR	Base activity  BA	Acid interaction  AI	HILIC retention & selectivity
Pore Size 		Hydrophobic selectivity  HS	Chelation  C	Ion exchange capacity pH 2.6  IEX(2.6)	
Particle Size 		Steric selectivity  SS	Ion exchange capacity (pH 7.6)  IEX(7.6)		
Particle Size Distribution 		Hydrogen bonding capacity  HBC			

Column Testing

Every Accucore HPLC column is also tested after packing and again compared to specifications and issued with a certificate.

Retention time	Capacity factor	Efficiency	Asymmetry	Backpressure
				

This material and column testing is evidence that every Accucore HPLC column will give reproducible chromatography, batch after batch, column after column.

Long Lasting Columns


Chromatographers today demand long lifetimes from the columns they use.

Mechanical Stability and Stable Bonded Phase

The highly uniform packed bed in Accucore HPLC columns is created by the use of tightly controlled particle size and automated packing process and has excellent mechanical stability.

The advanced bonding technology used for Accucore HPLC columns creates robust bonded phases that are highly resistant to the effects of pH and temperature.

Accucore HPLC columns show excellent stability at pH <2




Column:	Accucore C18 2.6 μ m, 100 x 2.1mm														
Mobile phase:	A – Water + 0.1% Trifluoroacetic Acid B – Methanol + 0.1% Trifluoroacetic Acid														
Gradient:	<table><thead><tr><th>Time (min)</th><th>% B</th></tr></thead><tbody><tr><td>0</td><td>25</td></tr><tr><td>0.75</td><td>25</td></tr><tr><td>10.00</td><td>100</td></tr><tr><td>12.00</td><td>100</td></tr><tr><td>12.20</td><td>25</td></tr><tr><td>17.00</td><td>25</td></tr></tbody></table>	Time (min)	% B	0	25	0.75	25	10.00	100	12.00	100	12.20	25	17.00	25
Time (min)	% B														
0	25														
0.75	25														
10.00	100														
12.00	100														
12.20	25														
17.00	25														
Flow rate:	400 μ L/min														
Injection volume:	1 μ L														
Temperature:	30°C														
Detection:	UV at 254nm (0.1s rise time, 20Hz)														
Order of elution:	<ol style="list-style-type: none">1. Uracil (t_r)2. Acetaminophen3. p-Hydroxybenzoic acid4. O-Hydroxybenzoic acid5. Amitriptyline6. Nortriptyline7. Di-isopropyl phthalate8. Di-n-propyl phthalate														


*Accucore HPLC columns
are robust and long lasting.*




Accucore HPLC columns are also stable at pH >10




Column:	Accucore C18 2.6 μ m, 100 x 2.1mm	
Mobile phase:	A – Water + 0.1% Ammonia B – Methanol + 0.1% Ammonia	
Gradient:	Time (min)	% B
	0	15
	1.00	15
	8.00	100
	11.00	100
	11.20	15
	16.00	15
Flow rate:	400 μ L/min	
Injection volume:	1 μ L	
Temperature:	30°C	
Detection:	UV at 254nm (0.1s rise time, 20Hz)	
Order of elution:	<ol style="list-style-type: none"> 1. Uracil (t_r) 2. 4-Chlorocinnamic acid 3. Procainamide 4. 4-Pentylbenzoic Acid 5. N-Acetylprocainamide 6. Di-isopropyl phthalate 7. Di-n-propyl phthalate 	



And also stable at elevated temperature



Column:	Accucore C18 2.6 μ m, 100 x 2.1mm	
Mobile phase:	35:65 (v/v) Water/Methanol	
Flow rate:	400 μ L/min	
Injection volume:	1.5 μ L	
Temperature:	70°C	
Detection:	UV at 254nm (0.1s rise time, 20Hz)	
Order of elution:	<ol style="list-style-type: none"> 1. Theophylline/Caffeine (t_r) 2. Phenol 3. Butylbenzene 4. o-Terphenyl 5. Pentylbenzene/Triphenylene 	




Column Protection

Guard columns are designed to protect your column from particulates introduced from the matrix or instrument and from any strongly retained components in the injected sample.

Thermo Scientific Defender Guard Cartridges have been designed specifically to work with high speed, high efficiency separations.


Backpressure v Injection Number



Accucore C18 2.6 μ m, 100mm x 2.1mm
10mm x 2.1mm Defender Guard Cartridge

Mobile:	35:65 (v/v) Acetonitrile/Water
Flow:	550 μ L/min
Temperature:	40°C
Detection:	UV at 254nm
Injection volume:	1 μ L
Protein crash sample:	rat plasma diluted 1 in 4 with Acetonitrile
	1 injection test mix followed by 9 protein crash injections

Defender Guard Cartridges extend the lifetime of your Accucore HPLC column.



Accucore C18 2.6 μ m, 100mm x 2.1mm
10mm x 2.1mm Defender Guard Cartridge

Mobile phase:	50:50 (v/v) Acetonitrile/Water
Flow rate:	400 μ L/min
Temperature:	30°C
Detection:	UV at 254nm
Injection volume:	1 μ L
Analytes:	<ol style="list-style-type: none"> 1. Theophylline 2. p-Nitroaniline 3. Methyl Benzoate 4. Phenetole 5. o-Xylene

Defender Guard Cartridges have minimal effect on column performance.


Parameter	With Defender Guard	Without Defender Guard
Selectivity	1.81	1.81
Efficiency (N/m)	185672	213350
Tailing	1.18	1.07
Retention Time (minutes)	3.19	2.90
Backpressure (bar)	280	256



Optimum Selectivity


Accucore HPLC columns are available in a series of phases to provide a wide range of selectivities. Each of the bonded phases is manufactured using advanced bonding technology and is characterized using a testing regime based on the Tanaka Tests. See page 15 for further details of these tests.

The radar plots below show the results of the characterization and allow for quick and easy comparison of the phase selectivities.




Accucore RP-MS

Optimized for MS detection, excellent combination of speed and quality of separation




Accucore C18

Optimum retention for non-polar analytes




Accucore aQ

Compatible with 100% aqueous mobile phases, special selectivity for polar analytes



Phenyl-Hexyl

Unique selectivity for aromatic and moderately polar analytes



Accucore PFP



Alternative selectivity to C18, particularly for halogenated analytes

HILIC

Accucore HILIC

Enhanced retention of polar and hydrophilic analytes

Thermo Scientific Accucore RP-MS




- Optimized for MS detection
- Excellent Peak Shapes
- Excellent Combination of Speed and Efficiency

Accucore RP-MS uses an optimized alkyl chain length for more effective coverage of the silica surface. This coverage results in a significant reduction in non-hydrophobic interactions and thus highly efficient peaks with very low tailing.

RP-MS offers slightly lower retention than C18 and this combined with high efficiencies and low peak tailing make this the phase of choice for use with MS detection.

The selectivity offered by Accucore RP-MS matches that of C18 columns.

Bases



Accucore RP-MS 2.6µm, 50mm x 2.1mm

Mobile phase: 65:35 (v/v) Methanol/25mM Potassium Phosphate pH7.0

Flow rate: 500µL/min

Temperature: 30°C



Detection: UV at 215nm

Injection volume: 1µL

Back pressure: 232 bar

Analytes: 1. Uracil
2. Propranolol
3. Butylparaben
4. Naphthalene
5. Acenaphthene
6. Amitriptyline

Thermo Scientific Accucore C18




- Optimum retention of non-polar compounds
- Hydrophobic interaction mechanism
- Separates a broad range of analytes

The carbon loading of Accucore C18 phase provides high retention of non-polar analytes via a predominantly hydrophobic interaction mechanism.

The highly retentive nature of Accucore C18 phase means that it can be used to separate a broad range of analytes.

Triazines



Accucore C18 2.6µm, 50mm x 2.1mm

Mobile phase: A – Water ; B – Acetonitrile

Gradient:	Time (min)	% B
	1.0	35
	2.5	70

Flow rate: 600µL/min

Temperature: 25°C

Detection: UV at 280nm



Injection volume: 2µL

Backpressure: 298 bar

Analytes:

1. Simazine
2. Simetryn
3. Atrazine
4. Ametryn
5. Propazine
6. Prometryn

Thermo Scientific Accucore aQ




- Retention and Resolution of Polar Analytes
- Polar Endcapped C18 Stationary Phase for Alternative Selectivity
- Ideal for Highly Aqueous Mobile Phases

The polar functional group used to endcap Accucore aQ phase provides an additional controlled interaction mechanism by which polar compounds can be retained and resolved, making Accucore aQ phase ideal for the quantitative analysis of trace levels of polar analytes.



The wettability of reversed phase media can be increased by the introduction of polar functional groups. The polar endcapping of Accucore aQ media also makes it usable in 100% aqueous mobile phases without the risk of loss of performance or poor stability.

Lamivudine (USP)



Accucore aQ 2.6µm, 50mm x 2.1mm	
Mobile phase:	95:5 (v/v) Ammonium Acetate, pH 3.80 / Methanol
Flow rate:	200µL/min
Temperature:	35°C
Detection:	UV at 277nm
Injection volume:	1µL
Analytes:	Lamivudine
%RSD t_r	0.00
%RSD Peak area	1.72
(%RSD calculated from 6 replicate injections)	
USP acceptance criteria: % RSD (t_r, Peak Area) <2.0	

Thermo Scientific Accucore Phenyl-Hexyl




- Mixed-mode Selectivity for Aromatic and Moderately Polar Analytes
- Enhanced Pi-pi Interactions with Aromatics
- Moderate Hydrophobicity

The C6 chain in Accucore Phenyl-Hexyl phase exhibits classical RP retention and selectivity, while the phenyl ring can add special selectivity by interacting with polar groups within the solutes. This results in a mixed-mode separation mechanism.

Phenyl-Hexyl phase should be selected for complex samples where some peaks are well resolved on a conventional alkyl phases, but are not well resolved on a conventional phenyl phase. While other peaks are well resolved on a phenyl phase, but not well resolved on a conventional alkyl phase.



Beta-agonists



Accucore Phenyl-Hexyl 2.6µm, 100mm x 2.1mm

Mobile phase:	A – Ammonium acetate 5mM, pH 4 B – Acetonitrile	
Gradient:	Time (min)	% B
	0	5
	1	5
	10	100
Flow rate:	0.25mL/min	
Temperature:	40°C	
Detection:	+ESI-MS (45°C, 4.5kV, 60V, scan 150 – 350)	
Injection volume:	1µL	
Backpressure:	120 bar	

Thermo Scientific Accucore PFP




- Alternative Selectivity to C18
- Extra Retention for Halogenated Species
- Unique Selectivity for Non-halogenated Polar Compounds

Introduction of fluorine groups into the Accucore PFP (pentafluorophenyl) stationary phase causes significant changes in solute-stationary phase interactions. This can lead to extra retention and selectivity for positional isomers of halogenated compounds.

PFP Columns are also well suited to the selective analysis of non-halogenated compounds, in particular polar compounds containing hydroxyl, carboxyl, nitro, or other polar groups. High selectivity is often most apparent when the functional groups are located on an aromatic or other rigid ring system.

Positional Isomers



Accucore PFP 2.6µm, 50mm x 2.1mm

Mobile phase: A – Water + 0.1% Formic Acid
B – Acetonitrile + 0.1% Formic Acid

Gradient: 15 – 30% B in 7 minutes

Flow rate: 600µL/min


Temperature: 50°C

Detection: UV at 270nm

Injection volume: 2µL

1. 3,4 – Dimethoxyphenol
2. 2,6 – Dimethoxyphenol
3. 2,6 – Difluorophenol
4. 3,5 – Dimethoxyphenol
5. 2,4 – Difluorophenol
6. 2,3 – Difluorophenol
7. 3,4 – Difluorophenol
8. 3,5 – Dimethylphenol
9. 2,6 – Dimethylphenol
10. 2,6 – Dichlorophenol
11. 4 – Chloro-3-Methylphenol
12. 4 – Chloro-2-Methylphenol
13. 3,4 – Dichlorophenol
14. 3,5 – Dichlorophenol

Thermo Scientific Accucore HILIC



HILIC


Particle Size 2.6µm **USP** L3

- Enhanced Retention of Polar and Hydrophilic Analytes
- Alternative Selectivity to C18 without Ion-pair or Derivatization
- Improved Sensitivity for MS detection

Analyte properties that govern retention with Accucore HILIC phase are acidity/basicity, which determines hydrogen bonding, and polarizability which determines dipole-dipole interactions.

The highly organic mobile phases used with Accucore HILIC phase ensure efficient desolvation in ESI MS detection, which in turn leads to improved sensitivity.

Catecholamines



Accucore HILIC 2.6µm, 150 x 4.6mm

Mobile phase: 85:15 (v/v) Acetonitrile/100mM Ammonium Formate, pH 3.2

Flow rate: 2mL/min

Temperature: 40°C

Detection: UV at 280nm

Injection volume: 5µL


Backpressure: 157 bar

Analytes:
1. Catechol
2. 5-HIAA
3. DOPAC
4. Serotonin
5. L-tyrosine
6. Dopamine
7. L-DOPA

Applications

Food

Azo Dyes



Accucore RP-MS 2.6 μ m, 150mm x 4.6mm

Mobile phase: A – Water + 0.1% Trifluoroacetic Acid
B – Acetonitrile + 0.1% Trifluoroacetic Acid

Gradient:	Time (min)	% B
	0	25
	3	30
	15	95
	17	95

Flow rate: 1.8mL/min


Temperature: 50°C

Detection: UV at 225nm

Injection volume: 2 μ L

Analytes: 1. Fast Garnet
2. Orange II
3. Dimethyl Yellow
4. Sudan Red G
5. Sudan I
6. Sudan II
7. Sudan IV

Melamine and Cyanuric Acid



Accucore HILIC 2.6 μ m, 150mm x 4.6mm

Mobile phase: 90:10 (v/v) Acetonitrile:50mM Ammonium Acetate, pH 5

Flow rate: 1mL/min

Temperature: 40°C

Detection: MS at m/z 127, 128, 168 (negative mode 0-3 mins, positive mode 3-10 mins)

Injection volume: 5 μ L


Backpressure: 117 bar

Analytes: Cyanuric Acid: m/z 128.1 (-1)
Melamine: m/z 127.1 (+1), 168.1 (+1 with Acetonitrile adduct)



Environmental

Uron Herbicides



Accucore RP-MS 2.6 μ m, 100mm x 2.1mm

Mobile phase: A – Water; B – Acetonitrile

Gradient: 35 – 60% B in 3.20 min

Flow rate: 0.43mL/min


Temperature: 25°C

Detection: UV at 240nm

Injection volume: 3 μ L

Analytes: 1. Monuron
2. Metoxuron
3. Linuron
4. Diuron
5. Terbutiuron
6. Chlortoluron

Banned Aromatic Amines



Accucore PFP 2.6 μ m, 100mm x 2.1mm

Mobile phase: A – 25mM Ammonium Acetate pH 5.0
B – Acetonitrile

Gradient: 20 – 100% B in 1.5 minutes

Flow rate: 0.6mL/min

Temperature: 40°C

Detection: UV at 254nm

Injection volume: 1 μ L


Backpressure: 300 bar

Analytes: 1. 2,4 – Diaminotoluene
2. 4,4 – Oxydianiline
3. o-Toluidine
4. 2-Methoxy-5-methylaniline
5. 2,4,5-Trimethylaniline
6. 4,4-Methylene-bis(2-chloroaniline)

*Basic polar compounds, volatile buffer
in the mobile phase – good peak shape.*

Clinical

Epinephrine and Norepinephrine



Accucore HILIC 2.6 μ m, 150mm x 4.6mm

Mobile phase: 92:8 (v/v) Acetonitrile/100mM Ammonium Formate, pH 3.2

Flow rate: 3.5mL/min


Temperature: 40°C

Detection: UV at 230nm

Injection volume: 10 μ L

Backpressure: 250 bar

Analytes: 1. Epinephrine
2. Norepinephrine



Accucore aQ 2.6 μ m, 100mm x 2.1mm

Mobile phase: 5mM Ammonium Formate, pH3.2

Flow rate: 0.6mL/min

Temperature: 30°C


Detection: UV at 230nm

Injection volume: 2 μ L

Analytes: 1. Norepinephrine (0.33 min)
2. Epinephrine (0.37 min)

Good retention and resolution of the two analytes in HILIC compared with the aQ column in reversed phase mode.

Steroids



Accucore RP-MS 2.6 μ m, 100mm x 2.1mm

Mobile phase: 50:50 (v/v) Water / Acetonitrile

Flow rate: 0.6mL/min


Temperature: 40°C

Detection: UV at 220nm

Injection volume: 1 μ L

Analytes: 1. Estradiol
2. Estrone

Speed – Sub 1 minute separation.



Accucore RP-MS 2.6 μ m, 100mm x 2.1mm

Mobile phase: A – Water; B – Acetonitrile

Gradient: 25 – 70% B in 4 minutes

Flow rate: 0.6mL/min

Temperature: 25°C

Detection: UV at 254nm


Injection volume: 1 μ L

Analytes: 1. Hydrocortisone
2. Cortisone
3. Corticosterone
4. 11-alpha Hydroxyprogesterone
5. 17-alpha Hydroxyprogesterone
6. Progesterone



Forensic Toxicology


Testosterones




Accucore RP-MS 2.6 μ m, 100mm x 2.1mm
Mobile phase: 60:40 (v/v) Water/Acetonitrile
Flow rate: 0.6mL/min
Temperature: 40°C
Detection: UV at 254nm
Injection volume: 1 μ L
Analytes: 1. 11-Ketotestosterone 2. 19-Nortestosterone (Nandrolone) 3. Epitestosterone

Forensics

Explosives




Accucore PFP 2.6 μ m, 100mm x 2.1mm
Accucore C18 2.6 μ m, 100mm x 2.1mm
Mobile phase: 60:40 (v/v) Water/Methanol
Flow rate: 0.4mL/min
Temperature: 40°C
Detection: UV 254nm
Injection volume: 1 μ L
Backpressure: 260 bar
Analytes: 1. 1,3,5-trinitrobenzene 2. Nitrobenzene 3. 1,3-dinitrobenzene 4. 2,4,6-trinitrotoluene 5. 2,4-dinitrotoluene



The carbon-fluorine bond on the PFP bonded phase is more polar than the carbon-hydrogen bond in the C18 phase, which results in extra retention and selectivity of these polar substituted aromatic compounds containing nitro functional groups.

Pharmaceuticals

Acidic and Neutral Drugs



Accucore C18 2.6 μ m, 100mm x 2.1mm

Mobile phase: A – Water + 0.1% Phosphoric Acid
B – Acetonitrile + 0.1% Phosphoric Acid

Gradient:	Min	% B
	0	30
	2	40
	4	95

Flow rate: 0.4mL/min

Temperature: 25°C

Detection: UV at 215nm


Injection volume: 1.5 μ L

Analytes:

1. Hydrochlorothiazide
2. Prednisolone
3. Pravastatin
4. Carbamazepine
5. Diclofenac
6. Ibuprofen
7. Progesterone

Excellent resolution.

Zidovudine (USP)



Accucore C18 2.6 μ m, 100mm x 2.1mm

Mobile phase: 80:20 (v/v) Water/Methanol

Flow rate: 0.5mL/min

Temperature: 25°C

Detection: UV at 265nm


Injection volume: 1 μ L

Analytes:	1. Zidovudine	USP specification
Asymmetry	1.13	< 2.0
%RSD t_r	0.0	< 2.0
%RSD Peak area	0.34	< 2.0



The Accucore C18 column demonstrated excellent performance for the analysis of Zidovudine. The USP criteria was exceeded.

Ibuprofen



Accucore C18 2.6 μ m, 100mm x 4.6mm

Mobile phase: 66.3% Water adjusted to pH 2.5 with Phosphoric Acid / 33.7% Acetonitrile

Flow rate: 2.0mL/min

Temperature: 30°C

Detection: UV at 214nm


Injection volume: 5 μ L

Back pressure: 276 bar

Analytes: 1. Valerophenone (USP grade)
2. Ibuprofen (USP grade)

The Accucore C18 column demonstrated excellent performance for the analysis of Ibuprofen, using a method based on the USP.

Fenoprofen and Gemfibrozil (USP)



Accucore RP-MS 2.6 μ m, 100mm x 2.1mm

Mobile phase: 50:49.6:0.4 (v/v/v) Acetonitrile/Water/Phosphoric Acid

Flow rate: 0.4mL/min

Temperature: 30°C


Detection: UV at 272nm

Injection volume: 3 μ L

Analytes: 1. Fenoprofen 2. Gemfibrozil USP specification

Asymmetry	1.23	1.22	< 2
%RSD t_r	0.39	0.39	< 2
%RSD Peak area	0.39	0.00	< 2
Resolution			> 8
Efficiency	7464	8259	> 3000

Lisinopril



Accucore RP-MS 2.6 μ m, 100mm x 2.1mm

Mobile Phase: A – 0.026M Monobasic Sodium Phosphate pH 3.75 with Phosphoric Acid
B – 80:20 (v/v) Mobile Phase A/Acetonitrile

Gradient:	Time (min)	% B
	0	3
	15	30
	30	30
	35	3

Flow rate: 0.4mL/min

Temperature: 50°C

Detection: UV at 210nm

Injection volume: 5 μ L

Backpressure: 159 bar


Analytes: 1. Impurity 1
2. 2-amino-4-Phenylbutyric Acid
3. Lisinopril
4. Lisinopril R,S,S isomer

Analytes:	Lisinopril	USP specification
Resolution 2,1	4.7	> 1.5
Resolution 2,3	12.7	> 1.5
Resolution 4,3	4.7	> 2.0
%RSD Peak	1.74	< 5%
Area Lisinopril		

The Accucore RP-MS column demonstrated excellent performance for the analysis of Lisinopril. The USP criteria was exceeded.

Bio-pharmaceuticals

Peptide Test Mix



Accucore C18 2.6 μ m, 100mm x 2.1mm

Mobile phase: A – Water + 0.1% Trifluoroacetic Acid
B – Acetonitrile + 0.1% Trifluoroacetic Acid

Gradient: 10 – 70 % B in 6 min

Flow rate: 0.50mL/min

Temperature: 40°C

Detection: UV at 220nm


Injection volume: 2 μ L

Analytes: 1. Glycyl-L-Tyrosine (MW=238.3)
2. Val-Tyr-Val (MW=379.5)
3. Methionine Enkephalin (MW=573.7)
4. Leucine Enkephalin (MW=555.6)
5. Ribonuclease A (MW=13700)
6. Insulin (MW=5733.5)

The Accucore C18 column can be used to analyze small peptides; demonstrated up to MW 13700.

General

High Throughput Ketones



Accucore C18 2.6µm, 50mm x 2.1mm

Mobile phase: A – Water; B – Acetonitrile

Gradient: Min % B

0 40

0.4 95

0.80 95

Flow rate: 1.0mL /min


Temperature: 45°C

Detection: UV at $\lambda = 258\text{nm}$

Injection volume: 1µL

Backpressure: 113 bar

Analytes: 1. 2-Pentanone
2. Acetophenone
3. 2-Heptanone
4. Butyrophenone
5. Hexanophenone
6. Octanophenone
7. Decanophenone



fully porous C18 5µm, 150mm x 4.6mm

Mobile phase: A – Water; B – Acetonitrile

Gradient: Min % B

0.00 40

2.30 95

4.60 95

Flow rate: 1.0 mL/min

Temperature: 45°C

Detection: UV at $\lambda = 258\text{nm}$

Injection volume: 10µL

Pressure: 47 bar

Analytes: 1. 2-Pentanone
2. Acetophenone
3. 2-Heptanone
4. Butyrophenone
5. Hexanophenone
6. Octanophenone
7. Decanophenone

Separation with the Accucore HPLC column is 6 times faster than the 150 x 4.6mm, 5µm fully porous column.

Thermo Scientific Options for Fast HPLC

Uniquely in the market Thermo Scientific HPLC columns offer chromatographers three different options for fast HPLC.

Hypersil GOLD 1.9 μ m and Synchronis 1.7 μ m – Small Particle Advantage

When used with UHPLC systems Hypersil GOLD 1.9 μ m and Synchronis 1.7 μ m HPLC columns offer optimum efficiency over a range of flow rates.

This high efficiency results in higher throughput, lower costs and better separations.

Accucore – Core Enhanced Technology

The core enhanced technology used in Accucore HPLC columns provides better speed, sensitivity and resolving power than 5 μ m and 3 μ m columns without the higher backpressures associated with sub-2 μ m materials.

Accucore HPLC columns do not require UHPLC systems.

Hypercarb – High Temperature HPLC

Hypercarb is a porous graphitized carbon material suitable for use at temperatures up to 200°C. As a result it is uniquely suited for use with high temperature HPLC.

High temperatures in HPLC increase efficiency, reduce backpressure and reduce analysis time.



Ordering Information

Accucore HPLC Columns

Description	Particle Size	Length (mm)	2.1mm ID	3.0mm ID	4.6mm ID
Accucore RP-MS	2.6µm	30	17626-032130	17626-033030	17626-034630
		50	17626-052130	17626-053030	17626-054630
		100	17626-102130	17626-103030	17626-104630
		150	17626-152130	17626-153030	17626-154630
Accucore C18	2.6µm	30	17126-032130	17126-033030	17126-034630
		50	17126-052130	17126-053030	17126-054630
		100	17126-102130	17126-103030	17126-104630
		150	17126-152130	17126-153030	17126-154630
Accucore aQ	2.6µm	30	17326-032130	17326-033030	17326-034630
		50	17326-052130	17326-053030	17326-054630
		100	17326-102130	17326-103030	17326-104630
		150	17326-152130	17326-153030	17326-154630
Accucore Phenyl-Hexyl	2.6µm	30	17926-032130	17926-033030	17926-034630
		50	17926-052130	17926-053030	17926-054630
		100	17926-102130	17926-103030	17926-104630
		150	17926-152130	17926-153030	17926-154630
Accucore PFP	2.6µm	30	17426-032130	17426-033030	17426-034630
		50	17426-052130	17426-053030	17426-054630
		100	17426-102130	17426-103030	17426-104630
		150	17426-152130	17426-153030	17426-154630
Accucore HILIC	2.6µm	30	17526-032130	17526-033030	17526-034630
		50	17526-052130	17526-053030	17526-054630
		100	17526-102130	17526-103030	17526-104630
		150	17526-152130	17526-153030	17526-154630

Accucore Defender Guard Columns (4/pk)

Description	Particle Size	Length (mm)	2.1mm ID	3.0mm ID	4.6mm ID
Accucore RP-MS	2.6µm	10	17626-012105	17626-013005	17626-014005
Accucore C18	2.6µm	10	17126-012105	17126-013005	17126-014005
Accucore aQ	2.6µm	10	17326-012105	17326-013005	17326-014005
Accucore Phenyl-Hexyl	2.6µm	10	17926-012105	17926-013005	17926-014005
Accucore PFP	2.6µm	10	17426-012105	17426-013005	17426-014005
Accucore HILIC	2.6µm	10	17526-012105	17526-013001	17526-014005

UNIGUARD Direct-Connection Guard Cartridge Holders

Description	2.1mm ID	3.0mm ID	4.6mm ID
UNIGUARD Drop-In Guard Cartridge Holder	852-00	852-00	850-00
Standard Replacement Tip	850-RT	850-RT	850-RT

Accucore Kits

Accucore kits allow validation of the reproducibility of Accucore HPLC columns, or verification of which Accucore HPLC column offers the best separation.

Accucore Validation Kit

Validate the reproducibility of Accucore HPLC columns. Contains 3 Accucore C18 columns.

Description	Particle Size	Length (mm)	2.1mm ID
Accucore Validation Kit	2.6µm	50	17126-052130-3V
		100	17126-102130-3V
		150	17126-152130-3V

Accucore Narrow Selectivity Kit

Verify which Accucore HPLC column offers the best separation over a narrow selectivity range. Contains 1 each of Accucore C18, RP-MS and aQ columns.

Description	Particle Size	Length (mm)	2.1mm ID
Accucore Narrow Selectivity Kit	2.6µm	50	17X26-052130-3VA
		100	17X26-102130-3VA
		150	17X26-152130-3VA

Accucore Wide Selectivity Kit

Verify which Accucore HPLC column offers the best separation over a wide selectivity range. Contains 1 each of Accucore C18, Phenyl-Hexyl and PFP columns.

Description	Particle Size	Length (mm)	2.1mm ID
Accucore Wide Selectivity Kit	2.6µm	50	17X26-052130-3VB
		100	17X26-102130-3VB
		150	17X26-152130-3VB

Accucore Polar Selectivity Kit

Verify which Accucore HPLC column offers the best separation for polar analytes. Contains 1 each of Accucore aQ, PFP and HILIC columns.

Description	Particle Size	Length (mm)	2.1mm ID
Accucore Polar Selectivity Kit	2.6µm	50	17X26-052130-3VC
		100	17X26-102130-3VC
		150	17X26-152130-3VC



Resources

for Chromatographers

Thermo Scientific Chromatography Columns and Consumables Catalog

This extensive catalog offers 540 pages of proven chromatography tools and product selection guides. Available online, with a robust search tool and optimized for your iPad®.

Visit www.thermoscientific.com/catalog



Chromatography Resource Center

Our web-based resource center provides technical support, applications, technical tips and literature to help move your separations forward.

Visit www.thermoscientific.com/chromatography



For more information visit: www.thermoscientific.com/accucore

© 2011 Thermo Fisher Scientific Inc. All rights reserved. iPad is a registered trademark of Apple Inc., registered in the US and other countries. Agilent is a registered trademark of Agilent Technologies Inc. All other trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

BRCCSACCUCORE 0611

USA and Canada +1 800 332 3331
France +33 (0)1 60 92 48 34
Germany +49 (0) 2423 9431 -20 or -21
Switzerland +41 56 618 41 11
United Kingdom +44 1928 534110
Japan +81 3 5826 1615

China +86-21-68654588 or +86-10-84193588
1 800-810-5118
India 1 800 22 8374 (toll-free) or +91 22 6716 2200
Australia 1300 735 292 (free call domestic)
New Zealand 0800 933 966 (free call domestic)
All Other Enquiries +44 (0) 1928 534 050

Technical Support
North America +1 800 332 3331
Outside North America +44 (0) 1928 534 440

Thermo
SCIENTIFIC

Part of Thermo Fisher Scientific