

**Gas Control Center**

Datei Produkte Benutzer Messdaten Extra

NEW

Test\_IP\_MAPY mapy\_192,168.6.120

PROD_NAME	PROD_NO	BARCODE	CHARGE	USER	MEAS.TIME	TIME	DATE	STATUS	INDEX	INFO	GAS_O2	GAS_CO2	GAS_N2	GAS_F
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 15:38:53	09.05.2008	0	19	O.K	1.71	23.8	74.5	974	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 15:39:18	09.05.2008	0	20	O.K	1.11	24.6	74.3	963	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:02:10	09.05.2008	0	21	O.K	0.61	25.4	73.8	971	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:02:36	09.05.2008	0	22	O.K	1.21	25.0	73.8	967	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:03:03	09.05.2008	0	23	O.K	1.21	24.7	74.1	962	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:31:57	09.05.2008	0	24	O.K	1.01	25.0	74.0	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:32:20	09.05.2008	0	25	O.K	1.61	24.5	73.9	978	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 16:32:43	09.05.2008	0	26	O.K	1.01	25.2	73.8	966	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:01:35	09.05.2008	0	27	O.K	-0.50	25.3	74.2	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:02:01	09.05.2008	0	28	O.K	1.71	24.3	74.0	970	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:02:32	09.05.2008	0	29	O.K	1.01	24.4	74.6	965	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:25:22	09.05.2008	0	30	O.K	0.71	25.2	74.1	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:25:56	09.05.2008	0	31	O.K	1.61	23.8	74.4	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 17:26:19	09.05.2008	0	32	O.K	1.41	24.2	74.4	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:23:00	09.05.2008	0	33	O.K	0.60	25.5	73.9	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:23:26	09.05.2008	0	34	O.K	1.41	24.5	74.1	964	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:23:51	09.05.2008	0	35	O.K	1.01	24.5	74.5	969	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:46:21	09.05.2008	0	36	O.K	0.91	24.9	74.2	966	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:46:45	09.05.2008	0	37	O.K	1.62	23.6	74.6	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 18:47:10	09.05.2008	0	38	O.K	1.01	24.5	74.5	970	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:19:57	09.05.2008	0	39	O.K	1.01	25.1	73.9	968	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:20:21	09.05.2008	0	40	O.K	1.31	24.6	74.1	971	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:20:49	09.05.2008	0	41	O.K	1.71	23.9	74.4	971	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:56:39	09.05.2008	0	42	O.K	1.21	24.6	74.2	967	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:57:04	09.05.2008	0	43	O.K	1.71	24.2	74.1	965	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 19:57:29	09.05.2008	0	44	O.K	1.31	24.3	74.4	970	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 20:23:54	09.05.2008	0	45	O.K	2.02	23.6	74.4	971	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 20:24:21	09.05.2008	0	46	O.K	1.51	24.2	74.3	975	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 20:24:53	09.05.2008	0	47	O.K	1.31	23.9	74.8	969	
TEST automatisch	1234567	124567	1	ADMINISTRATOR	12 21:17:38	09.05.2008	0	48	O.K	1.71	23.8	74.5	969	

G A S C O N T R O L

**Windows-Software for the documentation of analysis-results of OXYBABY®6.0 /-Med, MAPY, LEAK-MASTER®, MAPMAX and other WITT- equipment with integrated or separate analysis.**

**Benefits**

- quality control by means of data transfer and storing
  - date and time
  - user (name or examiner)
  - lot data
  - product line data
- export data interface for further interpretation of measured data in MS EXCEL®
  - data transfer to Control Unit (touch screen control)
  - product data
  - administration of product names
  - administration of user names inclusive access code
- logging and readable from up to 21 analysers
- multilingual interface: German, English, Spanish, Italian, Polish, French, Dutch (more to follow)

**System requirements**

- Operating system**      WINDOWS® 2000, XP, Vista, 7
- Hardware**                graphic resolution min. 800 x 600 pixel  
 standard Ethernet cable installed by customer
- Compatibility**            all OXYBABY®6.0 /-Med, MAPY, LEAK-MASTER®, MAPMAX  
 and gas mixers with integrated analysers  
 – please specify device number with order –

Product Information / Technical Data

B01/B1 subject to change