



Wittlogger - 2006-08-10_Daten_01.wlm

Artikelnummer: <Auto> Charge: 001

Artikelbezeichnung: <Auto> Anmerkung: ...

angelegt von / Prüfer: PR-001 Geräte: OXYBABY V2 (Gerät1)

Meßdaten download

Nr.	Gerät	Gas	Messwert	Alarm	Text	Datum	Zeit	Artikelnr.	Artikelbez.	Charge	Anmerkung	Prüfer
12	OXYBABY V2	O2	20,0			10.10.06	11:11:53	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,6			10.10.06	11:11:53	0002	Prod. 02	001	...	PR-001
13	OXYBABY V2	O2	20,0			10.10.06	11:12:12	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,6			10.10.06	11:12:12	0002	Prod. 02	001	...	PR-001
14	OXYBABY V2	O2	18,4	MinAlarm		10.10.06	11:12:26	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	1,3	MaxAlarm		10.10.06	11:12:26	0002	Prod. 02	001	...	PR-001
15	OXYBABY V2	O2	19,9			10.10.06	11:20:00	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,4	MinAlarm		10.10.06	11:20:00	0002	Prod. 02	001	...	PR-001
16	OXYBABY V2	O2	20,0			10.10.06	11:20:24	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,5			10.10.06	11:20:24	0002	Prod. 02	001	...	PR-001
17	OXYBABY V2	O2	20,1			10.10.06	11:20:40	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,4	MinAlarm		10.10.06	11:20:40	0002	Prod. 02	001	...	PR-001
18	OXYBABY V2	O2	19,7			10.10.06	11:20:58	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	1,4	MaxAlarm		10.10.06	11:20:58	0002	Prod. 02	001	...	PR-001
19	OXYBABY V2	O2	19,7			10.10.06	11:21:40	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,5			10.10.06	11:21:40	0002	Prod. 02	001	...	PR-001
20	OXYBABY V2	O2	19,9			10.10.06	11:21:59	0002	Prod. 02	001	...	PR-001
	OXYBABY V2	CO2	0,5			10.10.06	11:21:59	0002	Prod. 02	001	...	PR-001

Gesendet: 22 Empfangen: 38

1: Aktiv COM1 2: ... 3: ... 4: ... 5: ... 6: ... 7: ... 8: ... www.wittgas.com

Windows-Software for the documentation of analysis-results of OXYBABY®6.0, MFA, PA and other WITT- equipment with control system (AWS).

Benefits

- quality control by means of data recording and storing with date and time
- individual product administration of lot no., article no., examiner etc.
- data-export interface for further read-out of the measured data in MS EXCEL®
- simultaneous logging from up to 8 analysers
- multilingual interface
- automatic identification of the connected WITT-Analyser
- administration of product names

System requirements

- Operating system** WINDOWS® 2000, XP, Vista, 7
- Hardware** 1 free serial interface, graphic resolution min. 800 x 600 Pixel
- Compatibility** all MFA, PA, OXYBABY®6.0, WITT-analysers with AWS please specify device number with order
- Accessories** USB connecting cable included in delivery
- Option** RS232-adaptor cable

Product Information

Technical Data