



TJ Williams Ltd

Wimborne Road,
Barry
Vale of Glamorgan
CF63 3DH
Tel: 01446 729200
Fax: 01446 739281

VAT Reg. No. GB135255384
Co. Reg. No. 03611520

Web: www.tjwilliams.co.uk Email: sales@tjwilliams.co.uk

Manufacturers of Level, Pressure, and Temperature Instrumentation

CERTIFICATE OF CALIBRATION

CUSTOMER NAME:	CUSTOMER ORDER No.	292805
----------------	--------------------	--------

PRODUCT DETAILS

CERT TYPE	Definitive	CERTIFICATE No.	372476
PRODUCT	Pressure Gauge	DATE	21/12/2017
SERIAL No.	372476	ACCURACY +/- FS	1.60%
DIAL RANGE	0 to 400 psi & 28 bar	ACCURACY UNITS	6.40
TEST RANGE 0 /	400	REC. CAL. FREQ'Y	1 Year
TEST UNITS	psi	SALES O/No.	372476
SPECIFICATION	EN837	CAL. MEDIUM	Water

RISING RESULTS

FALLING RESULTS

PRODUCT READING	STANDARD READING	CORRECTION ERROR	STANDARD READING	CORRECTION ERROR
psi	psi	psi	psi	psi
0	0	0.00	0	0.00
100	100	0.00	100	0.00
200	201	1.00	200	0.00
300	301	1.00	300	0.00
350	349	-1.00	349	-1.00
400	396	-4.00		

The above product has been tested and found to conform to the specifications, unless stated below.

COMMENTS:-

DETAILS OF CALIBRATION EQUIPMENT USED

STANDARD	Test Gauge	Primary Standard	DWT
SERIAL No.	927471-3	UKAS LAB No.	0138
ACCURACY	0.25%	CERT No.	13241/13240
EXPIRY DATE	21/03/2017		

The calibration equipment used is within its period of calibration and is traceable to UKAS/International Standards.

Signed

Authorised Signature